

EZMeta™ α -Amylase Colorimetric Activity Kit

Cat #: D-AKC1370

Size: 96T

Storage: Stored at 4°C for 12 months, protected from light

Product Information

Detection range: 0.0156-1 mg/mL

Sensitivity: 0.0078 mg/mL

Applicable samples: Serum, Plasma, Saliva, Animal and Plant Tissues (seeds or newly germinated seedlings)

Assay Principle

Starch hydrolase includes α -amylase and β -amylase. α -AL (EC 3.2.1.1) randomly catalyzes the hydrolysis of α -1,4-glycosidic bonds in starch to generate reducing sugars such as glucose, maltose, maltotriose, dextrin, etc., while reducing the viscosity of starch, so it is also called liquefaction enzyme. α -amylase is widely present, from microorganisms to advanced plant. EZMeta™ α -Amylase Colorimetric Activity Kit provides a convenient tool for sensitive detection of α -AL activity. The principle is that starch hydrolase catalyzes the hydrolysis of starch to form reducing sugars, and 3,5-dinitrosalicylic acid is reduced by reducing sugar to produce a brown-red substance with an absorption peak at 540 nm; the amylase activity is calculated by measuring the rate of increase in absorbance at 540 nm. α -AL is heat-resistant, but β -amylase can be inactivated at 70°C for 15 min. Therefore, only α -AL can catalyze starch hydrolysis after the crude enzyme solution is passivated at 70°C for 15 min.

Materials Supplied and Storage Conditions

Kit components	Size	Storage conditions
	96 T	
DNS Reagent	40 mL	4°C, protected from light
Substrate	1	4°C
Standard	1	4°C

Materials Required but Not Supplied

- Microplate reader or visible spectrophotometer capable of measuring absorbance at 540 nm
- 96-well plate or microglass cuvette, precision pipettes, disposable pipette tips
- Centrifuge, water bath
- Deionized water
- Homogenizer (for tissue samples)

Reagent Preparation

DNS Reagent: Ready to use as supplied. Equilibrate to room temperature before use. Store at 4°C, protected from light.

Substrate: Add 20 mL deionized water before use, then shake upside down several times and heat to dissolve. This solution can be stored at 4°C. If there is precipitation, it can be heated at 70°C to dissolve.

Standard: Add 1 mL of deionized water to dissolve the standard substance (glucose) at 10 mg/mL before use. which could be stored at 4°C for 1 month. The solution can also be stored at -20°C for long time.

Standard curve setting: Dilute 10 mg/mL standard with deionized water to 1, 0.5, 0.25, 0.125, 0.0625, 0.0313, 0.0156 mg/mL standard solution as shown in the table below.

Num.	Volume of Standard (μL)	Volume of Deionized Water (μL)	The Concentration of Standard (mg/mL)
Std.1	40 μL (10 mg/mL)	360	1
Std.2	200 μL of Std.1 (1 mg/mL)	200	0.5
Std.3	200 μL of Std.2 (0.5 mg/mL)	200	0.25
Std.4	200 μL of Std.3 (0.25 mg/mL)	200	0.125
Std.5	200 μL of Std.4 (0.125 mg/mL)	200	0.0625
Std.6	200 μL of Std.5 (0.0625 mg/mL)	200	0.0313
Std.7	200 μL of Std.6 (0.0313 mg/mL)	200	0.0156

Sample Preparation

Note: Fresh samples are recommended.

1. Animal tissues: Weigh 0.1 g tissue, add 1 mL deionized water and homogenize. Pour the homogenate into a centrifuge tube, leave it at room temperature for 15 min, and shake it every 5 min to make it fully extracted. Centrifuge at 6,000 g for 10 min at room temperature. Aspirate the supernatant and add deionized water to make the volume up to 10 mL. Shake well to obtain the amylase stock solution.

2. Plant tissues: Weigh 0.1 g tissue, add 1 mL deionized water and mash. Ultrasonic break 5 min (power 20%, ultrasonic 3 s, interval 7 s, repeat 30 times). Leave it at room temperature for 15 min, and shake it every 5 min to make it fully extracted. Centrifuge at 6,000 g for 10 min at room temperature. Aspirate the supernatant and add deionized water to make the volume up to 10 mL. Shake well to obtain the amylase stock solution.

3. Liquid samples such as plasma, serum and saliva: Tested directly. It is recommended to determine the appropriate dilution factor in a pre-experiment.

Note: For animal tissues with high fat content, remove the upper layer of fat after centrifugation, and then take the supernatant. It will be better to quantify the total protein with EZMeta™ α-Amylase Colorimetric Activity Kit, if the content is calculated by protein concentration.

Assay Procedure

1. Preheated the microplate reader or visible spectrophotometer for more than 30 min, and adjust the wavelength to 540 nm, visible spectrophotometer was returned to zero with deionized water.
2. Preheat the water bath to 70°C.
3. 75 μ L of samples were placed in a boiling water bath for 5 min and use it as a control tube.
4. Add the following reagents respectively into each EP tube:

Reagent	Blank Tube (μ L)	Standard Tube (μ L)	Test Tube (μ L)	control Tube (μ L)
Deionized Water	75	0	0	0
Stds.	0	75	0	0
Sample	0	0	75	75 (boiled sample)

Placed in a 70°C water bath for 15 min and allow to cool

Substrate	0	0	75	0
-----------	---	---	----	---

Keep warm for 5 min in a 40°C water bath

DNS Reagent	150	150	150	150
Substrate	75	75	0	75

5. Mix well, placed in a boiling water bath for 5 min and allow to cool. Take out 200 μ L to a 96-well plate or microglass cuvette. Then reading the values at 540 nm. Finally, calculate $\Delta A_{\text{Test}} = A_{\text{Test}} - A_{\text{Control}}$, $\Delta A_{\text{Standard}} = A_{\text{Standard}} - A_{\text{Blank}}$.

Note: Each sample has a control tube, only one blank well needs to be detected. In order to guarantee the accuracy of experimental results, need to do a pre-experiment with 2-3 samples. If the A_{Test} is higher than 2, please further dilute the sample with deionized water. Pay attention to multiply by the dilution factor when calculating the result. If the ΔA_{Test} is less than 0.005, the sample can be re-extracted and reduce the volume of deionized water.

Data Analysis

1. Drawing of standard curve

With the concentration of the standard solution as the y-axis and the $\Delta A_{\text{Standard}}$ as the x-axis, draw the standard curve.

Substitute the ΔA_{Test} into the equation to obtain the y value (mg/mL).

2. Calculate the activity of α -Amylase in sample

(1) Calculated by fresh weight of samples

Unit definition: 1 mg reducing sugar produced per min in 1 g tissue reaction system is defined as a unit of enzyme activity.

$$\alpha\text{-Amylase (U/g weight)} = y \times V_{\text{Sample}} \div (W \times V_{\text{Sample}} \div V_{\text{Total Sample}}) \div T \times n = 2 \times y \div W \times n$$

(2) By protein concentration

Unit definition: 1 mg reducing sugar produced per min in 1 mg tissue protein reaction system is defined as a unit of enzyme activity.

$$\alpha\text{-Amylase (U/mg prot)} = y \times V_{\text{Sample}} \div (Cpr \times V_{\text{Sample}}) \div T \times n = 0.2 \times y \div Cpr \times n$$

(3) Calculated by volume of liquid sample

Unit definition: 1 mg reducing sugar produced per min in 1 L liquid sample reaction system is defined as a unit of enzyme activity. α -Amylase (U/L) = $1,000 \times y \div T \times n = 200 \times y \times n$

Where: V_{Sample} : sample volume added, 0.075 mL; W: sample weight, g; $V_{\text{Total Sample}}$: Total volume of sample, 10 mL; T: reaction time, 5 min; n: dilution factor; Cpr: sample protein concentration, mg/mL; 1000: 1 L = 1000 mL.

Typical Data

Typical standard curve:




Figure1. Standard curve for glucose.

Disclaimer

The reagent is only used in the field of scientific research, not suitable for clinical diagnosis or other purposes.