

EZMeta™ Serum Potassium (K⁺) Colorimetric Assay Kit

Cat #: D-AKC2100

Size: 96T

Storage: Stored at 4°C for 12 months, protected from light

Product Information

Applicable samples: Serum

Assay Principle

Potassium can maintain the normal osmotic pressure and acid-base balance of the body, participate in glucose metabolism and protein metabolism. EZMeta™ Serum Potassium (K⁺) Colorimetric Assay Kit provides a simple method for detecting K⁺ content in serum samples. The resulting turbidity is proportional to the K⁺ concentration within a certain range. Serum K⁺ content was determined by measuring turbidity.

Materials Supplied and Storage Conditions

Kit components	Size	Storage conditions
	96 T	
Extraction Buffer	50 mL	4°C
Reagent I	2 mL	4°C
Reagent II	1	4°C, protected from light
Reagent III	20 mL	4°C
Standard	1 mL	4°C

Materials Required but Not Supplied

- Microplate reader or visible spectrophotometer capable of measuring absorbance at 520 nm
- 96-well plate or microglass cuvette, precision pipettes, disposable pipette tips
- Centrifuge
- Deionized water

Reagent Preparation

Extraction Buffer: Ready to use as supplied; Equilibrate to room temperature before use; Store at 4°C.

Working Reagent II : Prepare before use, Take Reagent I , add all to Reagent II , then mixing. Store at 4°C, protected from light.

ReagentIII: Warm to 25°C in incubator for more than 30 min before use.

Standard: Ready to use as supplied. Equilibrate to room temperature before use. Store at 4°C.

Sample Preparation

Serum samples: Add 50 µL Serum, 450 µL Extraction Buffer into the EP tube. Mix thoroughly. Centrifuge at 8,000 rpm for 10 min at 25°C and aspirating the supernatant to be tested.

Assay Procedure

1. Preheat the microplate reader or visible spectrophotometer for more than 30 min, and adjust the wavelength to 520 nm, visible spectrophotometer was returned to zero with deionized water.
2. Sample measurement (the following operations are operated in the 96-well plate or microglass cuvette).

Reagent	Blank Well (μL)	Standard Well (μL)	Test Well (μL)
Deionized Water	40	0	0
Standard	0	40	0
Supernatant	0	0	40
Working Reagent II	20	20	20
Mix well and kept it for 5 min.			
ReagentIII	140	140	140

3. Mix well, the absorbance value is measured at 520 nm. The blank well is marked as A_{Blank} , the standard well is marked as A_{Standard} , and the test well is marked as A_{Test} .

Data Analysis

Note: We provide you with calculation formulae, including the derivation process and final formula. The two are exactly equal. It is suggested that the concise calculation formula in bold is final formula.

$$K^+ \text{ (mmol/L)} = [C_{\text{Standard}} \times (A_{\text{Test}} - A_{\text{Blank}}) \div (A_{\text{Standard}} - A_{\text{Blank}})] \times n = \mathbf{5 \times (A_{\text{Test}} - A_{\text{Blank}}) \div (A_{\text{Standard}} - A_{\text{Blank}})}$$

C_{Standard} : 0.5 mmol/L; n : Sample dilution factor, 10.

Disclaimer

The reagent is only used in the field of scientific research, not suitable for clinical diagnosis or other purposes.